

## EU DECLARATION OF CONFORMITY

**Manufacturer:** GRI Medical & Electronic Technology Co., Ltd.

1805 Honggao Road  
314031 Jiaxing, Zhejiang  
PEOPLE'S REPUBLIC OF CHINA

**MF's SRN:** CN-MF-000004203

**EU Authorized Representative:** Emergo Europe B.V.

Westervoortsedijk 60,  
6827 AT Arnhem,  
The Netherlands

**AR's SRN:** NL-AR-000000116

**Device:** Securement Device (refer to see the attached 1)

**Basic UDI-DI:** 69319181F005KT

**Intended purpose:** The purpose of securement devices is to secure cables and wires away from the surgical site within the surgical field to prevent contamination of the surgical site. It is used sterile.

**Classification:** Class I sterile, according to MDR 2017/745 Annex VIII, Rule 1

**Notified Body Name:** TUV SUD


**Notified Body Identification Number:** 0123

**Conformity Assessment Procedure:** Annex IX Chapters I and III, MDR 2017/745

**CE Certificate:** G11 056820 0042 Rev.02

**Applicable standards or Common Specifications:** N/A

The device covered by the present EU declaration is in conformity with the (EU) MDR 2017/745. All supporting documentations are retained under the premises of the manufacturer. We are exclusively responsible for the declaration of conformity.

**Signature/Date:**  2024/12/23

**Print Name/Position:** Sidonia Dong VP RA&QA Asia Pacific

**Place:** Jiaxing, China

Attached 1

<b>Part #</b>	<b>Description</b>
25-5751-S	Tube positioner_Sterile_50ea/cs
25-5751C-S	Tube positioner_Sterile_30ea/cs
AR-30802	Tube positioner_Sterile_50ea/cs
25-6750-S	Tubing & Cable Manager_6.75" x 2" _Sterile_400ea/cs